

Proglider files – next step in glide path preparation

HC Baranwal¹, Akash Kumar Baranwal^{2,*}

¹Associate Professor, ²Senior Resident, Dept. of Conservative Dentistry & Endodontics, Faculty of Dental Sciences, IMS, BHU, Varanasi, Uttar Pradesh

***Corresponding Author:**

Akash Kumar Baranwal

Senior Resident, Dept. of Conservative Dentistry & Endodontics, Faculty of Dental Sciences, IMS, BHU, Varanasi, Uttar Pradesh

Email: baranwalakash@yahoo.com

Abstract

As the field of Endodontics has become revolutionized with development of rotary Ni-Ti files, the advancements of these file systems have improved the cleaning & shaping of root canal with time. Before using these Ni-Ti rotary instruments, a glide path preparation within canal seems to be very necessary. The endodontic glide path is a smooth radicular tunnel from canal orifice to physiologic apical terminus (apical foramina constriction). Glide Path can be generated either with stainless steel file system or rotary Ni-Ti Files. Recently, M -wire technology has been incorporated in endodontic file systems, which offers greater flexibility and resistance against cyclic fatigue. So, the present review article introduces the Proglider Glide path File (Dentsply mallifer) which is the next step in glide path management with progressive taper and M-Wire metallurgy, securing canals that have never been easier or safer.

Keywords: Glide Path, Rotary instrument, M-Wire technology, Proglider, Root canal treatment.

Introduction

One of the mechanical objectives of root canal instrumentation is to achieve a continuous tapering funnel shape of the original canal from the coronal access to the apex. To achieve this goal, a glide path within canal must be discovered or prepared. The glide path can be short or long, narrow or wide, essentially straight or curved. Ni-Ti instrument when used directly in the root canal and starts rotating, they may experience structural fatigue which eventually lead to failure.¹ Torsion and fatigue through flexure are the main two reasons for rotary Ni-Ti instrument separation within canal.^{2,3} The flexural fatigue (bending stress) resulting in to fracture usually occurs when an instrument (already weakened by metal fatigue) is kept under stress and the amount of such stress transforming onto a file depends on the root canal anatomy especially the canal curvature.⁴ Usually, torsional fracture occurs when the tip or any other part of the rotating instrument bind to the root canal walls while rest of the file keeps turning.⁵ The reasons for torsional fracture may include exerting too much apical force on the rotating instrument in the root canal or when there is a wide area of contact between the cutting flutes of the instrument and the canal wall or when the canal section is smaller than the dimension of the non-cutting or non-active tip of the instrument.⁶ Considering the flexural fatigue, the instrument does not bind to the canal walls, but rotate freely until the separation of instrument occurs, at the point of maximum flexure.⁷ Factors associated with Ni-Ti rotary instruments separation may include – instrument design, instrumentation technique, rotational speed, and torque on instrument. To minimize the risk

of instrument fracture, performing coronal enlargement of root canal becomes very essential.⁸

Perhaps the greatest obstacle performing endodontic treatment is to consistently find, follow, and secure any given canal from coronal to its apical terminus. For this, Glide path management requires flawless performance which may be defined as the ability to execute a task in a prescribed manner. Completely negotiating and securing canal with small sized hand files requires a mechanical skilful touch, patience, and desire. A small- sized hand file is initially used to scout, expand, and refine the internal wall of the canal. Once the canal can be manually reproduced, a mechanical glide path file may be used to expand the working width in preparation for shaping procedures.

The glide path should be created with small flexible stainless steel manual file to verify within any portion of root canal. This will allow sufficient space for the rotary instrument to follow. Proglider files which come in single file system, generates a faster & safer glide path preparation. Proglider file is made of M-wire Ni-Ti alloy to provide greater flexibility and resistance against cyclic fatigue. Before understanding the concepts of such path files, the M-wire technology has to be known.

M-wire Technology

M-wire (Sportwire LLC, Langley, OK, USA) is a new Ni-Ti wire that has been developed through proprietary thermo-mechanical processing. M-wire has significantly improved fatigue resistance of endodontic rotary instruments in comparison with those made of conventional superelastic Ni-Ti alloy.^{10,11,12} M-wire consists of 508 nitinol, which undergo a processing

method comprising of thermo-mechanical treatment of the raw wire under specific tensile stresses and temperature.¹⁰

Fine microstructural differences between the two wires can be seen under Transmission Electron Microscope (TEM) imaging of non-deformed wire. The Ni-Ti wire consisted essentially the β phase austenite with a dislocation-cell substructure and precipitate whereas the M-wire microstructure contained martensite, with cell substructure and precipitates.¹³

Features of Pro Glider File:¹⁴

Proglider is single Ni-Ti file for glide path preparation. It is most advanced glide path solution in terms efficiency, simplicity and safety. Features are –

- One Single mechanical glide path file
- Progressive tapers from 2% to 8% over its length
- Offers a smoother 'glide path' Transition
- Utilising M-wire technology
- Controlled, smooth, inward cutting action
- Continuous rotation, 300rpm/2ncm, adjustable up to 5.2ncm
- Three Blade Length - 21mm, 25mm, 31mm.
- The Pro glider file has a diameter of 0.16mm at D0 and 0.82 mm at D 16.

Proglider file is suitable for most root canals including the highly curved ones. Pro glider File also respect the root canal anatomy better than manual glide path file, which make it difficult to avoid the risk of canal transportation, zips and ledges.

Discussion

The mechanical objective of biomechanical preparation is to get a continuous tapering funnel shape of root canal from coronal access to the apex. To achieve this goal, a glide path within canal must be prepared. There are four skills that require glide path preparation. First skill is to find the canal, second is to follow the canal, third skill is to understand the four possible reasons that you may not follow the radiographic terminus and last technique is to understand and master the four manual motions to prepare the rotary glide path. Four reasons for a file not following to its radiographic terminus, include when canal is blocked, file curvature does not replicate canal curvature, diameter of file is too wide at its tip and diameter of file is too wide in its shafts. Sometimes, a combination of reason one to four, or all four may be the situation.⁹ Pro Glider path file is solution of all four problems. It removes the coronal dentin restriction due to progressive taper from 0.16mm to tip to 0.82mm at D16. Due to its high flexibility, Proglider respects the root canal anatomy better than manual glide path file.

Generally, during the endodontic therapy, Ni-Ti instruments are used under rotational speed ranging from 300 to 500 rpm, applied by an endodontic motor operating with torque varying from 3 to 5 N cm in a

16:1 reduction.^{15,16} Figueiredo et al¹⁷ showed that the number of cycle to failure in martensitic Ni-Ti wire (Proglider) can be as much as 100 times larger than in stable and superelastic austenitic Ni-Ti. M-Wire technology has improved fatigue behaviour due to presence of thin martensite variants. The presence of these martensite variants, which can be related to high transformation temperature found in M-wire.^{18,19}

According to Miyazaki and Otsuka²⁰, martensite variants can undergo stress – induced reorientation at relatively lower stress and this can be reason for the lower apparent elastic modulus of M-wire and smaller force required for bending by the same amount M-wire in relation to conventional Ni-Ti.

Alpati et al¹⁸ evaluated the first metallurgical characterization of M-wire and showed that under certain processing condition, M-wire consisted of austenite, martensite and R phase with their relative proportions depending on the processing conditions. They also reported that M –wire had higher transformation temperature, compared with a conventionally treated Ni-Ti wire employed for the manufacture of Ni-Ti instrument. Also, M-wire had a lower apparent elastic modulus as well as smaller transformation stress and mechanical hysteresis. So, the Proglider having M-wire technology showed superior physical and mechanical properties than conventional Ni-Ti wire for manufacturing rotary endodontic instrument.¹³

Conclusion

Preparing the Glide path with M-wire path file (pro Glider) has superior physical and mechanical properties than conventional Ni-Ti path file. It contains martensite in non-deformed microstructure and this is the probably the reason why M-wire (pro glider) had higher fatigue resistance under 6% strain level than the Ni-Ti wire. Due to progressive taper and offset design, Proglider Path Files follow & expand original anatomy, improve shaping results, reduce chair time and decrease post-operative pain.

References

1. Schilder H. Cleaning and shaping the root canal. Dent Clin North Am 1974;18:269-96.
2. West J. The endodontic glidepath: secret to rotary success. Dentistry Today 2010; 29(9):86,8,90-9343.
3. Serene TP, Adams JD, Saxena A. Nickel titanium instrument: application in endodontics. J Endod 1995;2:92-94.
4. Pruett JP, Clement DJ, Carnes DL Jr. Cyclic fatigue testing of nickel-titanium endodontic instrument. J Endod 1997;36:93-99.
5. Kobayashi C, Yoshioka T, Suda H. A new engine driven canal preparation system with electronic canal measuring capabilities. J Endod 1997;23:751-54.
6. Blum JY, Machtou P, Ruddell CJ, Micallef JP. The analysis of mechanical preparation in extracted teeth using protaper rotary instrument: value of safety quotient. J Endod 2003;29:567-75.

7. Sattapan B, Palmara JE, Messar HH. Torque during canal instrumentation using nickel-titanium files. *J Endod* 2000;26:156-60.
8. Peters OA, Peter CI, Schonenberg K, Barbakow F. Protaper rotary root canal preparation: assessment of torque and force in relation to canal anatomy. *Int Endod J* 2003;36:93-99.
9. West JD. Perforation, blocks, ledges and transportation: overcoming barriers to endodontic finishing. *Dent Today* 2005;24:68-73.
10. Johnson E, Lloyd A, Kuttler S, Namerow K. Comparison between a novel nickel- titanium alloy and 508 nitinol on the cyclic fatigue life of Profile 25/.04 rotary instrument. *J Endod* 2008;34:1406-9.
11. Al-Hadlaq SMS, Aljarbou FA, AiThumairy RI. Evaluation of cyclic flexural fatigue of M-Wire nickel titanium rotary instruments. *J Endod* 2010;36:305-7.
12. Gao Y, Shotton V, Wilkinson K, Phillips G, Johnson WB. Effect of raw material and rotational speed on the cyclic fatigue of profile Vortex rotary instrument. *J Endod* 2010;36:1205-9.
13. Pereira ESJ, Gomes RO, Leroy AMF, Singh RP, Peters OA, Bahia MGA et al. Mechanical behaviour of M-Wire and conventional NiTi wire used to manufacture rotary endodontic instruments. *Dent Mater* 2013;29: e318-24.
14. Ruddle CJ, Machtou P, West JD. Endodontic canal preparation: innovations in glide path management and shaping canals. *Dent Today* 2014;33(7):118.
15. Viera EP, Franca EC, Martins RC, Buono VTL, Bahia MGA. Influence of multiple clinical use in fatigue resistance of Protaper rotary nickel titanium root canal instruments. *Int Endod J* 2008;41:163-72.
16. Gao Y, Gutmann JL, Wilkinson K, Maxwell R Ammon D. Evaluation of the impact of raw materials on the fatigue and mechanical properties of Profile Vortex rotary instruments. *J Endod* 2012;38:398-401.
17. Figueiredo AM, Modenesi P, Buono V. Low cycle fatigue life of superelastic NiTi Wires. *Int J Fatigue* 2009;31:751-58.
18. Alpati SB, Brantley WA, Lijima M, Clark WA, Kowarik L, Buie C et al. Metallurgical characterization of a new nickel-titanium wire for rotary endodontic instruments. *J Endod* 2009;35:1589-93.
19. Pareira ESJ, Peixoto IFC, Viana ACD, Oliviera H, Gongalez BM, Buono VTL et al. Physical and mechanical properties of a thermomechanically treated NiTi wire used in the manufacture of rotary endodontic instruments, *Int Endod J* 2012;45:469-74.
20. Miyazaki S, Ostuka k. Deformation and transition behaviour associated with the R-phase in NiTi alloy. *Metallurg Mater Trans A* 1986;17(1):53-63.